

Original Article

Cite this article: Bridge P, Kirby MC, and Callender JA. (2020) Evaluating VERT as a radiotherapy plan evaluation tool: comparison with treatment planning software. *Journal of Radiotherapy in Practice* 19: 210–214. doi: [10.1017/S1460396919000797](https://doi.org/10.1017/S1460396919000797)

Received: 21 August 2019
Revised: 27 September 2019
Accepted: 28 September 2019
First published online: 25 October 2019

Key words:

education; evaluation; radiotherapy; treatment planning; VERT

Author for correspondence:

Dr Pete Bridge, University of Liverpool, Brownlow Hill, Liverpool L69 3GB, UK. Tel: +44(0)1517958366. E-mail: pete.bridge@liverpool.ac.uk

Evaluating VERT as a radiotherapy plan evaluation tool: comparison with treatment planning software

P. Bridge , M. C. Kirby  and J. A. Callender 

School of Health Sciences, University of Liverpool, Liverpool L69 3GB, UK

Abstract

Introduction: The virtual environment for radiotherapy training (VERT) helps students to gain technical skills and understanding of 3D anatomy and dosimetry. It has potential as a tool for treatment plan evaluation, although little formal evidence currently supports this.

Aim: This paper reports findings from a plan evaluation workshop that facilitated comparison of VERT plan evaluation tools with those provided by conventional treatment planning software (TPS).

Method: Students on a pre-registration Post-Graduate Diploma in Radiotherapy worked in small groups evaluating lung plans using both VERT and Eclipse TPS tools. All students were invited to provide ratings concerning how helpful each modality was for a range of evaluation parameters and preferences for use.

Results: Most students (11 out of 14) found the session useful and expressed a desire to use VERT in future plan evaluation. The TPS was perceived to be more helpful with constraint-based evaluation while VERT was more helpful with evaluating plans for clinical set-up and delivery ($p < 0.001$).

Conclusion: Student therapeutic radiographers found VERT to be helpful as a plan evaluation tool alongside standard TPS tools, in particular for clinical set-up and delivery aspects of planning. Future work is ongoing to identify the specific impact of VERT as a plan evaluation tool for both students and qualified planners.

Introduction

Radiotherapy planning practical experience is an integral aspect of pre-registration training. The knowledge and skills necessary to produce a clinically acceptable plan are vital preparation for both clinical treatment planning and delivery, especially for more complex, dynamic and adaptive techniques. In addition, from an educational perspective, treatment planning offers a useful format for integrating student understanding of anatomy, oncology, technique and radiobiology as well as instilling holistic patient-focused practice, despite planning and treatment sometimes being viewed in clinical practice as separate entities. Aside from practical skills assessment, students are frequently assessed on their ability to evaluate a radiotherapy plan against accepted dose targets and constraints. Students are able to utilise the tools provided by conventional treatment planning software (TPS) including dose–volume histograms (DVH), conformity Index¹ and automated planning metric reports.²

Since its introduction to radiotherapy education in 2007,³ the virtual environment for radiotherapy training (VERT) 3D visualisation platform has become an increasingly useful teaching resource for therapeutic radiography students,^{4–10} medical physics students,^{11,12} radiotherapy staff¹³ and also radiotherapy patients.¹⁴ This simulation software is used around the world^{15,16} to enable students to gain technical skills in a safe environment and to visualise patient anatomy, contoured volumes and dose distribution in large-screen 3D.⁴ The use of VERT as an aid to teaching treatment plan comparison and evaluation was discussed in a recent publication⁵ but as yet there has been no published data relating to this. This project aimed to evaluate the potential role of VERT in a radiotherapy plan evaluation workshop through a comparison with the conventional tools provided by a leading TPS solution.

Methods

All 24 students on a pre-registration Post-Graduate Diploma in Radiotherapy course attended a 3-hour workshop that provided them with plan evaluation experience. The aim of the session was to provide them with useful feedback that they could utilise in their summative assessments. The workshop presented students with three radical lung plans for the same patient dataset as their assessment. The plans comprised a conventional conformal plan, a static gantry intensity-modulated radiotherapy plan and a volumetric modulated arc therapy plan for

© The Author(s) 2019. Published by Cambridge University Press. This is an Open Access article, distributed under the terms of the Creative Commons Attribution licence (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted re-use, distribution, and reproduction in any medium, provided the original work is properly cited.

Table 1. Relative usefulness of evaluation modalities

	Usefulness								
	All students			ECLIPSE FIRST			VERT FIRST		
	ECLIPSE	VERT	Delta	ECLIPSE	VERT	Delta	ECLIPSE	VERT	Delta
Target objectives	8.6	5.1	3.5	8.4	6.1	2.3	8.7	4.1	4.6
Lung constraints	8.2	4.6	3.6	7.9	6.0	1.9	8.6	3.3	5.3
Heart constraints	8.2	5.2	3.0	8.0	6.0	2.0	8.4	4.2	4.2
Oesophagus constraints	8.3	5.5	2.8	8.1	6.0	2.1	8.4	5.0	3.4
Cord constraints	8.3	5.6	2.7	8.1	6.0	2.1	8.4	5.1	3.3
Ease of set-up	5.1	7.4	-2.3	5.7	8.0	-2.3	4.6	6.9	-2.3
Ease of delivery	5.1	7.5	-2.4	5.9	7.9	-2.0	4.4	7.1	-2.7
Overall mean	7.4	5.8	1.6	7.4	6.6	0.8	7.4	5.1	2.3

Bold values represent the majority response.

Table 2. Preferred formats of evaluation modalities

	Preferred format								
	All students			ECLIPSE FIRST			VERT FIRST		
	ECLIPSE	VERT	Mixture	ECLIPSE	VERT	Mixture	ECLIPSE	VERT	Mixture
Target objectives	6	0	8	2	0	5	4	0	3
Lung constraints	6	0	8	1	0	6	5	0	2
Heart constraints	6	0	8	1	0	6	5	0	2
Oesophagus constraints	5	1	8	1	0	6	4	1	2
Cord constraints	5	1	8	1	0	6	4	1	2
Ease of set-up	0	1	13	0	0	7	0	1	6
Ease of delivery	0	0	14	0	0	7	0	0	7

Bold values represent the majority response.

comparison. Students were split into groups of five, with an experienced tutor on hand for individual and group guidance. They were asked to use both the ECLIPSE TPS software (Varian Medical Systems, Palo Alto, CA, USA) and VERT version 3.2 (Vertual Ltd, Hull, England) to help with their plan evaluation and comparison. Each evaluation session took between 45 minutes and 1 hour with the order of evaluation tool randomised. All students had previously undertaken at least 20 hours of tutor-guided practical planning with Eclipse within two module assignments, yet had little experience of using VERT other than in treatment set-up simulation. Guidance was therefore provided via both tutor demonstration and provision of written information regarding the plan visualisation functions within VERT. This included interactive 3D visualisation of different machines, plans, dose distributions, contours and surfaces. Students were guided to display dose on orthogonal CT surfaces and experiment with different transparencies, dose levels and colour maps. They were also shown how to benefit from different viewpoints using the 3D navigation, pan, zoom and rotation functions. Beam visualisation and animation were demonstrated to help students to visualise delivery. In order to facilitate independent learning, students were encouraged to experiment with the software with tutors on hand to provide assistance if required.

After the session, all students were invited to provide feedback on their experience via an anonymous online survey (Survey Monkey TM). Rating questions used a 0–9 scale to gather data concerning how helpful each modality was for a range of evaluation

tasks and objectives as seen in Table 1. Additional Likert style questions sought feedback concerning preferences for use of each modality as seen in Table 2. Finally open questions encouraged further description of the perceived value of the technology and its use in radiotherapy planning education.

Rating responses were subjected to inferential statistical analysis with paired *t*-tests comparing whole cohort ratings of each modality. Independent *t*-tests also compared perceptions between groups using each modality in different orders. Descriptive statistics were used to summarise the Likert responses. Responses to open questions were analysed using thematic analysis with responses coded and collated into themes. Blind coding was performed by two independent researchers before themes were agreed.

University Research Ethics Committee approval was provided for this project. All students received information about the evaluation project and were advised that participation in data collection was voluntary and that all data were anonymous. It was also explained that participation status would not be known to the teaching team and would not affect student performance, support or opportunities. Informed consent was sought in relation to use of the survey data for evaluation purposes.

Results

Of the 24 students, 14 completed the online survey. Most students (13 out of 14) enjoyed the plan evaluation session and expressed a

Table 3. Comparison of means *t*-test results

Comparison	A	B	Test	Mean A	Mean B	<i>n</i>	<i>p</i>
Helpfulness (constraint evaluation)	ECLIPSE	VERT	Paired	8.3	5.2	70	<0.001
Helpfulness (ease of set-up evaluation)	ECLIPSE	VERT	Paired	5.1	7.5	28	<0.001
Perception of VERT helpfulness	ECLIPSE first	VERT first	Independent	6	4.3	70	0.001

Table 4. Most useful plan evaluation tools within VERT

Theme	Comments
Visualising delivery	Seeing the beam set-up and delivery
	Playing the treatment beams in situ
	Visualisation especially of set-up, delivery
	3D aspect: allows to see beam in 3D so you can evaluate by eye
	Looking at beam angles and assessing ease of delivery and set-up
	Being able to look at gantry positions, to assess ease of set-up
	Gantry rotation
	Gantry motion
	Seeing the gantry
Visualising dose	Visualisation of the plan
	Visualising dose to contralateral lung and low dose areas
	Visualisation of how the healthy lung is effected
	Showing where the dose is within organs
	Looking at coverage of PTV and amount of dose OAR tissue
Contouring—transparency tool to assess PTV	
Visualising anatomy	3D contours
	Transparency of organs
	Anatomical geography

desire to use VERT as an additional plan evaluation tool in the future. Most students (11 out of 14) found the session to be useful. Students were asked to rate the extent to which the two modalities helped them to evaluate their plans for a range of plan evaluation parameters. Table 1 shows the mean scores ranging from 0 to 9 for usefulness while Table 2 shows their choice of preferred format converted into '1's and '0's to indicate their choices.

Table 3 summarises the inferential analysis for the statistically significant data. Following testing for normality, for constraint evaluation, a paired *t*-test across all 14 students for all 5 constraint questions (70 datasets) demonstrated a mean increase of 3 points in favour of ECLIPSE in terms of helpfulness compared to VERT ($p < 0.001$). In addition, a paired *t*-test across the ease of set-up and delivery domains demonstrated a mean difference of 2.3 across 28 datasets (14 students comparing each modality) in favour of VERT usefulness when evaluating ease of set-up and delivery ($p < 0.001$).

An independent *t*-test was performed between the group that accessed VERT or ECLIPSE first to identify any differences in their perception of the usefulness of VERT dependent upon the order of

Table 5. Desired additional plan evaluation tools for VERT

Theme	Comments
DVH	DVHs
	DVHs
	No DVHs
	DVHs for dose constraints
	DVH values
Constraints	Being able to see over 100% for constraints such as D105
	Constraint values
	How to calculate dose accurately not just visualise
Control	Touch screen remote controller

Table 6. Desired additional plan evaluation tools for ECLIPSE

Theme	Comments
Visualising delivery	Gantry motion (x2)
	Being able to see the gantry
	Being able to look at the position of the gantry
	Being able to assess ease of delivery
Visualising dose and anatomy	More visual aids
	More visualisation
Control	Touch screen

evaluation. This showed a statistically significant difference ($p = 0.001$) in student scores of usefulness for evaluation of constraints in favour of VERT for the group that used ECLIPSE first (mean score of 6) compared to those who used VERT first (mean score of 4.3). There was no statistically significant difference in perception of ECLIPSE usefulness between the groups.

Student comments were collated into themes relating to which tools within VERT they found the most useful (Table 4) and what additional tool or functionality would have helped within VERT (Table 5) and ECLIPSE (Table 6). The final question challenged students to identify the role that VERT could play in plan evaluation as seen in Table 7.

Discussion

Role of VERT in plan evaluation

It was clear that overall the students perceived VERT to be a helpful tool for some key aspects of plan evaluation. The visualisation

Table 7. Future role of VERT in plan evaluation

Theme	Comments
Plan checking	Delivery checking
	To see how it would physically be delivered
	Helps with assessing ease of set-up
Teaching	Seeing whether the plan is feasible to deliver, whether the gantry will hit the bed or not
	Assistance knowing what exactly we should be looking for
	More use of VERT maybe when doing the planning lessons
Planning	Before treatment planning session
	Shows the organs with 3D dose and where it hits—cannot see specifically in Eclipse
	Visualisation of the plan and volumes and understanding of gantry movement
	Allows a rough glimpse and evaluation of constraints. However, should be checked on Eclipse for actual coverage

aspects in particular helped students to evaluate clinical delivery and set-up factors that could impact on plan viability. Comments indicated that being able to see the actual machine deliver the plan helped students to understand clinical delivery issues. Most of the students recommended that VERT should play a role in plan evaluation. Table 7 includes comments identifying the value of VERT for evaluating set-up and delivery where it was seen as a particularly important addition to TPS-based evaluation. These findings confirmed published predictions^{4,5} concerning the potential value of VERT for plan evaluation. Interestingly, the students that used Eclipse first as an evaluation tool all stated that their preferred method for evaluating each dose constraint or objective was to use both systems together, whereas the majority of those students who used VERT first suggested that Eclipse alone was their preferred format for assessment. In rating the value of each modality for assessing individual dose constraints, the value placed on Eclipse was similar for both groups while there was a higher reported value of VERT from the group using it second. It was clear that, despite the relative inexperience of the students with VERT, they had all managed to access the necessary functions. This suggests that the software functionality is intuitive and training requirements minimal. It was also interesting to note that VERT was perceived to be useful when evaluating dosimetric factors such as target and organ at risk (OAR) doses which is primarily the remit of a TPS, although the mean value for this functionality overall was significantly lower overall than when using Eclipse ($p = 0.001$). There was a clear acknowledgement, however, that while VERT helped to provide a useful overview of the plan and potential issues, a TPS was essential for formal plan evaluation. When students were asked what additional features would have made VERT more useful for this task, DVH depiction was a common request. This would certainly have provided more insight into the formal achievement of dose–volume constraints and target objectives.¹⁷

Educational role of VERT

In addition to helping with the plan evaluation exercise, students also suggested that VERT would have provided useful insight during their treatment planning teaching sessions. The visualisation of dose to OAR structures using intuitive mouse controls to change perspective, rotate and zoom was felt to be particularly useful, as

seen in Table 4. The ability to visualise the 3D dose and volume relationship in VERT provided visual feedback on choice of beam angles, dose homogeneity and volumes of over or under-dosage. In this study, the students learned about the impact of these factors on plan viability and were able to include this in their assessed written evaluations. It was interesting to see students discover additional useful functionality in VERT, for example visualising all beams at once on the VERT plan. It would be instructive to repeat this exercise with a larger sample of students as an interim plan evaluation tool and to measure what changes, if any, are made to plans as a result of visualisation in VERT. Inclusion of quantitative analysis of performance would also provide useful insight into the specific impact of VERT on plan evaluation. The findings from this study will be implemented locally with VERT-based plan evaluation being embedded in the curriculum for all relevant cohorts.

Limitations

The small scale of the study should be acknowledged as a limitation and future planned collaborative work will aim to increase the sample size by including data from multiple institutions. It should also be acknowledged that these data were provided by student radiographers and that experienced planners would perhaps have a better instinctive understanding of 3D positioning of fields and dose deposition. It would be interesting to repeat this study with experienced planners to gain their perspective on the specific value of VERT for clinical plan evaluation. Conversely, it would also be valuable to compare the experience of first-time users provided with identical training and guidance.

Conclusions

This small study has shown that student radiographers found value in using VERT for plan evaluation alongside standard TPS tools. The ability to visualise structures, dose and beam delivery in 3D provided students with increased understanding of the clinical set-up and delivery aspects of planning. Comments from the students also suggested that VERT should be used more frequently throughout their planning modules to enhance their understanding of dosimetric principles and relational CT anatomy. Future work is ongoing to identify the specific impact of VERT as a plan evaluation tool for both students and qualified planners.

References

1. Feuvret L, Noël G, Mazeron JJ, Bey P. Conformity index: a review. *Int J Radiat Oncol Biol Phys* 2006; 64: 333–342.
2. Bridge P, Warren M, Pagett M. Use of planning metrics software for automated feedback to radiotherapy students. *J Radiother Pract* 2016; 15 (4): 385–391.
3. Bridge P, Appleyard R, Ward J, Phillips R, Beavis A. The development and evaluation of a virtual radiotherapy treatment machine using an immersive visualisation environment. *Comput Educa* 2007; 49 (2): 481–494.
4. Kane P. Simulation-based education: a narrative review of the use of VERT in radiation therapy education. *J Med Radiat Sci* 2018; 65 (2): 131–136.
5. Chamunyonga C, Burbery J, Caldwell P, Rutledge P, Fielding A, Crowe S. Utilising the virtual environment for radiotherapy training system to support undergraduate teaching of IMRT, VMAT, DCAT treatment planning, and QA concepts. *J Med Imag Radiat Sci* 2018; 49 (1): 31–38.
6. Boejen A, Beavis A, Nielsen K et al. Training of radiation therapists using a 3D virtual environment. *Radiother Oncol* 2007; 84: S275.
7. Phillips R, Ward JW, Page L et al. Virtual reality training for radiotherapy becomes a reality. *St Heal T* 2008; 132: 366–371.
8. Beavis A, Ward J. The development of a virtual reality dosimetry training platform for physics training. *Med Phys* 2012; 39 (6): 3969.
9. Kirby MC. Teaching radiotherapy physics concepts using simulation: experience with student radiographers in Liverpool, UK. *Med Phys Int* 2015; 3 (2): 87–93.
10. Kirby MC. The VERT physics environment for teaching radiotherapy physics concepts – update of four years' experience. *Med Phys Int* 2018; 6 (2): 247–254.
11. Jimenez Y, Ronn Hansen C, Juneja P, Thwaites DI. Successful implementation of virtual environment for radiotherapy training (VERT) in medical physics education: the University of Sydney's initial experience and recommendations. *Australas Phys Eng Sci Med* 2017; 40: 909–916.
12. Jimenez Y, Thwaites DI, Juneja P, Lewis SJ. Interprofessional education: evaluation of a radiation therapy and medical physics student simulation workshop. *J Med Radiat Sci* 2018; 65: 106–113.
13. James S, Dumbleton C. An evaluation of the utilisation of the virtual environment for radiotherapy training (VERT) in clinical radiotherapy centres across the UK. *Radiography* 2013; 19 (2): 142–150.
14. Stewart-Lord A, Brown M, Noor S, Cook J, Jallow O. The utilisation of virtual images in patient information giving sessions for prostate cancer patients prior to radiotherapy. *Radiography* 2016; 22: 269–273.
15. Stewart-Lord A. From education to research: a journey of utilising virtual training. *J Radiother Pract* 2016; 15 (1): 58–90.
16. Bridge P, Giles E, Williams A, Boejen A, Appleyard R, Kirby M. International audit of virtual environment for radiotherapy training usage. *J Radiother Pract* 2017; 16: 375–382.
17. ICRU Report 83. Prescribing, recording, and reporting intensity-modulated photon-beam therapy (IMRT). International Commission on Radiation Units and Measurements, Bethesda, MD, 2010.